boettger

Technical Data (PD)

Lid for Medicine Cup, Snap Closure					
REF	06-114-0*05				

1.	**Description / Specification**			
1.01	Description	Lid for Medicine Cup REF 06-103-010*, snap closure		
1.02	Dimension	See *Customer Drawing*		
1.03	Volume	N/A		
1.04	Material / Resin	Polyethylene (LDPE)		
	Grade of Resin	MAT130		
	Properties of Resin	See *www.boettger.com*		
1.05	Colour	White		
	Grade of Pigment	PIG125		
1.06	Hygiene Level	Standard hygienic		
1.07	Quality Control	Testing of attributive and variable characteristics in accordance with the valid specification		
1.08	Usage / Purpose	Snap closure of medicine cup		
1.09	Conformities / Regulations	See *Classification*		
1.10	Product Stewardship	See *Product Stewardship* of Grades of Resin/Pigment		
1.11	Storage	in closed original packaging, dry, at room temperature		
1.12	Transport	Dry, at room temperature, handle with care, do not throw or push cartons, do not stack pallets		
1.13	Other information	For single use only, recyclable		

2.	**Features**	
2.01	Basic features	-
2.02	Temperature range	-50°C to 60°C
2.03	Autoclavability	No
2.04	Centrifugation	N/A
2.05	Chemical Resistance	See *www.boettger.com*
2.06	Expiry	5 years from LOT (YYYY-MM)
2.07	Other information	-

3.	**Packaging**	**06-114-0005**	**06-114-0105**	
3.01	Pcs/Bag	750	75	
3.02	Pcs/Ctn	3750	3750	
3.03	Pcs/Pallet	180000	90000	
3.06	LOT / VP	WWYY (week, year of production) / XX (packaging)		
3.07	Other information	See *Measures*		

4.	**Other Information**	
4.01	Codes and Numbers	See *Codes and Numbers*
4.02	Symbols	See *Symbols*
4.03	Legal Manufacturer	Paul Boettger GmbH & Co. KG Miesleuthenweg 35-45 D-94249 Bodenmais
4.04	CH-REP	Swiss AR Services AG, Industriestr. 47, CH-6300 Zug CHRN-AR-20000807

Confidential: Information contained in this document or drawing is confidential and proprietory to Paul Boettger GmbH & Co. KG. This document may not be reproduced for any reason without written permission from Paul Boettger GmbH & Co. KG. All rights of design, invention and copyright are reserved.

Heruntergeladen von Gast (IP 216.73.217.10) am 24.11.2026 um 11:15 Uhr · sana.swiss

5.	Label of Sales Unit (Example) according to MDR 2017/745/EC



6.	Printing on Subpackaging (Example)

Lid Medicine Cup 75 pcs
REF 06-114-0105 CE
2025-11-20
Paul Boettger GmbH & Co. KG

* Placeholder for packaging alternatives

All informations without guarantee. Above mentioned products are manufactured according to company specifications and our certified Quality Management System corresponding to DIN EN ISO 13485. The quality check is carried out on a sample basis; an individual examination is not possible. Deficiencies parts may occur. We accept no liability for consequential damages.

Heruntergeladen von Gast (IP 216.73.217.10) am 24.11.2026 um 11:15 Uhr · sana.swiss