Add value.
Inspire trust.

TÜV SÜD Product Service GmbH· Ridlerstr. 65 · 80339 Munich · Germany

P. J. Dahlhausen & Co. GmbH
Adam-Riese-Straße 4
50996 Köln

| Your reference/letter of | Our reference/name | Tel. extension/Email | Fax extension | Date | Page |
|---|---|---|---|---|---|
| 15692 | 713347895, 713347896, 713347898 | medical_devices@tuvsud.com | / | 2024-09-25 | 1 of 8 |

**TÜV SÜD Product Service GmbH**
**Confirmation Letter**
**CL 015692 0508 Rev. 00**

**Reference: 713347895 | 713347896 | 713347898**

To whom it may concern,

**Confirmation of the status of a formal application, written agreement, and appropriate surveillance in the framework of Regulation EU 2023/607 amending Regulations (EU) 2017/745 (in the following referenced as MDR) as regards the transitional provisions for certain medical devices and in vitro diagnostic medical devices.**

With this letter TÜV SÜD Product Service GmbH, designated under MDR and identified by the number 0123 on NANDO, confirms that we have received a formal application in accordance with Section 4.3, first subparagraph of Annex VII of MDR and has signed a written agreement in accordance with Section 4.3, second subparagraph of Annex VII of MDR with the above stated manufacturer with the following SRN Number:

SRN Number: DE-MF-000006357

The devices covered by the formal application and the written agreement mentioned above are identified in the Tables below.

- Table 1 identifies the devices for which an MDR application has been received, written agreement concluded and for which TÜV SÜD Product Service GmbH is also responsible for appropriate surveillance of the corresponding devices under the applicable Directive.
- Table 2 identifies the devices for which an MDR application has been received and a written agreement concluded, but TÜV SÜD Product Service GmbH has not yet taken the responsibility for appropriate surveillance of the corresponding devices under the applicable Directive.



Heruntergeladen von Gast (IP 216.73.217.10) am 23.11.2026 um 16:21 Uhr · sana.swiss

If devices covered by certificates issued under Directive 90/385/EEC (AIMDD) or Directive 93/42/EEC (MDD) that expired after 26 May 2021 and before 20 March 2023, without having been withdrawn, this letter also confirms that

- the manufacturer signed the written agreement under MDR by the date of MDD/AIMDD certificate expiry; or
- provided evidence that a competent authority of a Member State had granted a derogation or exemption from the applicable conformity assessment procedure in accordance with Article 59(1) of MDR or Article 97(1) of the MDR respectively.

The transition timelines in accordance Article 120 (3) of MDR that apply to the devices covered by this letter, subject to the manufacturer's continued compliance to the other conditions specified in Article 120 (3c) of MDR, are shown below:

- 26 May 2026 for Class III custom-made implantable devices
- 31 December 2027 for Class III devices and Class IIb implantable devices (except sutures, staples, dental fillings, dental braces, tooth crowns, screws, wedges, plates, wires, pins, clips and connectors)
- 31 December 2028 for other Class IIb devices, Class IIa, Class I devices placed on the market in sterile condition, measuring function
- 31 December 2028 for devices not requiring the involvement of a notified body under MDD but requiring it under MDR (e.g., class I devices that qualify as re-usable surgical instruments)

We reserve the right to invoice any issuance, copies, amendments and / or changes of the confirmation letter according to effort.

For confirmation letter validity see www.tuvsud.com/ps-cert?q=cert:CL 015692 0508 Rev. 00

In case of inquiries please contact medical_devices@tuvsud.com.

On behalf of the Notified Body TÜV SÜD Product Service GmbH,
2024-09-25

TÜV SÜD Product Service GmbH
Medical and Health Services

Torsten Alt
Conformity Assessment Responsible (CARE)

TÜV SÜD Product Service GmbH
Medical and Health Services

Arianit Fazlija
Application Reviewer

Heruntergeladen von Gast (IP 216.73.217.10) am 23.11.2026 um 16:21 Uhr · sana.swiss

**Table 1: Devices covered by this letter and for which TÜV SÜD Product Service GmbH is also responsible for appropriate surveillance of the corresponding devices under the applicable Directive:**

| Device name or Basic UDI-DI (under MDR application) | MDR Device classification (as proposed by the manufacturer and verified during application review) | If the MDR device is a substitute device, identification of the corresponding MDD/AIMDD device | MDD/AIMDD Certificate Reference(s) of the devices under MDR application, and the NB Identification |
|---|---|---|---|
| **34342Neutralelektrode3D** | ☒ Class IIb / Class IIb implantable (exempted) | ☒ N/A | ☒ Certification as follows:<br>Certificate # G1 015692 0502 Rev. 01<br>NB# 0123 |
| **34342SillikonballonND** | ☒ Class IIb implantable (non-exempted) | ☒ N/A | ☒ Certification as follows:<br>Certificate # G1 015692 0502 Rev. 01<br>NB# 0123 |
| **34342AbsaugkatheterWU** | ☒ Class IIa | ☒ N/A | ☒ Certification as follows:<br>Certificate # G2 015692 0503 Rev. 00<br>NB# 0123 |
| **34342AbsaugsystemG5** | ☒ Class IIa | ☒ N/A | ☒ Certification as follows:<br>Certificate # G1 015692 0502 Rev.01<br>NB# 0123 |
| **34342ASchlauchSQV** | ☒ Class I devices in sterile condition | ☒ N/A | ☒ Certification as follows:<br>Certificate # G2S 015692 0504 Rev. 01<br>NB# 0123 |
| **34342BeatmungsbGF** | ☒ Class IIa | ☒ N/A | ☒ Certification as follows:<br>Certificate # G1 015692 0502 Rev. 01<br>NB# 0123 |
| **34342BeatmungsH9** | ☒ Class IIa | ☒ N/A | ☒ Certification as follows:<br>Certificate # G1 015692 0502 Rev. 01<br>NB# 0123 |
| **34342BezugLampeEZ** | ☒ Class I devices in sterile condition | ☒ N/A | ☒ Certification as follows:<br>Certificate # G2S 015692 0504 Rev. 01<br>NB# 0123 |
| **34342CSchuheSG3** | ☒ Class I devices in sterile condition | ☒ N/A | ☒ Certification as follows:<br>Certificate # G2S 015692 0504 Rev. 01<br>NB# 0123 |
| **34342DarmrohrXL** | ☒ Class IIa | ☒ N/A | ☒ Certification as follows:<br>Certificate # G1 015692 0502 Rev. 01<br>NB# 0123 |
| **34342DrainageUD** | ☒ Class IIa | ☒ N/A | ☒ Certification as follows:<br>Certificate # G1 015692 0502 Rev. 01<br>NB# 0123 |
| **34342Einnahmeglas7T** | ☒ Class I devices with measuring function | ☒ N/A | ☒ Certification as follows:<br>Certificate # G3M 015692 0506 Rev. 00<br>NB# 0123 |
| **34342ElektrodenreinigerLF** | ☒ Class I devices in sterile condition | ☒ N/A | ☒ Certification as follows:<br>Certificate # G2S 015692 0504 Rev. 01<br>NB# 0123 |

Heruntergeladen von Gast (IP 216.73.217.10) am 23.11.2026 um 16:21 Uhr · sana.swiss

| Device name or Basic UDI-DI (under MDR application) | MDR Device classification (as proposed by the manufacturer and verified during application review) | If the MDR device is a substitute device, identification of the corresponding MDD/AIMDD device | MDD/AIMDD Certificate Reference(s) of the devices under MDR application, and the NB Identification |
|---|---|---|---|
| **34342EmbolekkVQ** | ☒ Class IIa | ☒ N/A | ☒ Certification as follows:<br>Certificate # G1 015692 0502 Rev. 01<br>NB# 0123 |
| **34342Endotrachealtubus34** | ☒ Class IIa | ☒ N/A | ☒ Certification as follows:<br>Certificate # G1 015692 0502 Rev. 01<br>NB# 0123 |
| **34342Endotrachfuehrung89** | ☒ Class I devices in sterile condition | ☒ N/A | ☒ Certification as follows:<br>Certificate # G1 015692 0502 Rev. 01<br>NB# 0123 |
| **34342FadenMS4V** | ☒ Class I devices in sterile condition | ☒ N/A | ☒ Certification as follows:<br>Certificate # G2S 015692 0504 Rev. 01<br>NB# 0123 |
| **34342FadenziehSA3** | ☒ Class I devices in sterile condition | ☒ N/A | ☒ Certification as follows:<br>Certificate # G2S 015692 0504 Rev. 01<br>NB# 0123 |
| **34342Filter79** | ☒ Class IIa | ☒ N/A | ☒ Certification as follows:<br>Certificate # G1 015692 0502 Rev. 01<br>NB# 0123 |
| **34342FrazierHY** | ☒ Class IIa | ☒ N/A | ☒ Certification as follows:<br>Certificate # G1 015692 0502 Rev. 01<br>NB# 0123 |
| **34342GasinsufflationWF** | ☒ Class IIa | ☒ N/A | ☒ Certification as follows:<br>Certificate # G1 015692 0502 Rev. 01<br>NB# 0123 |
| **34342GasprobensQA** | ☒ Class IIa | ☒ N/A | ☒ Certification as follows:<br>Certificate # G1 015692 0502 Rev. 01<br>NB# 0123 |
| **34342GefaessA8** | ☒ Class IIa | ☒ N/A | ☒ Certification as follows:<br>Certificate # G1 015692 0502 Rev. 01<br>NB# 0123 |
| **34342GuedeltubeN4** | ☒ Class I devices in sterile condition | ☒ N/A | ☒ Certification as follows:<br>Certificate # G2S 015692 0504 Rev. 01<br>NB# 0123 |
| **34342HautklammergeraetCK** | ☒ Class IIa | ☒ N/A | ☒ Certification as follows:<br>Certificate # G2 015692 0503 Rev. 00<br>NB# 0123 |
| **34342HautstanzeY3** | ☒ Class IIa | ☒ N/A | ☒ Certification as follows:<br>Certificate # G1 015692 0502 Rev. 01<br>NB# 0123 |
| **34342InfusionsWX** | ☒ Class I devices in sterile condition | ☒ N/A | ☒ Certification as follows:<br>Certificate # G1 015692 0502 Rev. 01<br>NB# 0123 |

Heruntergeladen von Gast (IP 216.73.217.10) am 23.11.2026 um 16:21 Uhr · sana.swiss

| Device name or Basic UDI-DI (under MDR application) | MDR Device classification (as proposed by the manufacturer and verified during application review) | If the MDR device is a substitute device, identification of the corresponding MDD/AIMDD device | MDD/AIMDD Certificate Reference(s) of the devices under MDR application, and the NB Identification |
|---|---|---|---|
| **34342IrrigationsSTB** | ☒ Class I devices in sterile condition | ☒ N/A | ☒ Certification as follows: Certificate # G2S 015692 0504 Rev. 01 NB# 0123 |
| **34342KatheterapliHT** | ☒ Class IIa | ☒ N/A | ☒ Certification as follows: Certificate # G1 015692 0502 Rev. 01 NB# 0123 |
| **34342KlammerentMM** | ☒ Class I devices in sterile condition | ☒ N/A | ☒ Certification as follows: Certificate # G2S 015692 0504 Rev. 01 NB# 0123 |
| **34342KuenstlicheNase9P** | ☒ Class IIa | ☒ N/A | ☒ Certification as follows: Certificate # G1 015692 0502 Rev. 01 NB# 0123 |
| **34342KVerschlussPF** | ☒ Class I devices in sterile condition | ☒ N/A | ☒ Certification as follows: Certificate # G2S 015692 0504 Rev. 01 NB# 0123 |
| **34342LaryngealmaskeMT** | ☒ Class IIa | ☒ N/A | ☒ Certification as follows: Certificate # G1 015692 0502 Rev. 01 NB# 0123 |
| **34342LatexballonkHE** | ☒ Class IIa | ☒ N/A | ☒ Certification as follows: Certificate # G1 015692 0502 Rev. 01 NB# 0123 |
| **34342Nabelk3F** | ☒ Class I devices in sterile condition | ☒ N/A | ☒ Certification as follows: Certificate # G2S 015692 0504 Rev. 01 NB# 0123 |
| **34342OrthosaugerMG** | ☒ Class IIa | ☒ N/A | ☒ Certification as follows: Certificate # G1 015692 0502 Rev. 01 NB# 0123 |
| **34342PEEPVentilYB** | ☒ Class IIa | ☒ N/A | ☒ Certification as follows: Certificate # G1 015692 0502 Rev. 01 NB# 0123 |
| **34342RDrainage7M** | ☒ Class IIa | ☒ N/A | ☒ Certification as follows: Certificate # G1 015692 0502 Rev. 01 NB# 0123 |
| **34342Redon6P** | ☒ Class IIa | ☒ N/A | ☒ Certification as follows: Certificate # G1 015692 0502 Rev. 01 NB# 0123 |
| **34342Reservoir2J** | ☒ Class I devices in sterile condition | ☒ N/A | ☒ Certification as follows: Certificate # G2S 015692 0504 Rev. 01 NB# 0123 |
| **34342SauerstoffkF7** | ☒ Class IIa | ☒ N/A | ☒ Certification as follows: Certificate # G1 015692 0502 Rev. 01 NB# 0123 |

Heruntergeladen von Gast (IP 216.73.217.10) am 23.11.2026 um 16:21 Uhr · sana.swiss

| Device name or Basic UDI-DI (under MDR application) | MDR Device classification (as proposed by the manufacturer and verified during application review) | If the MDR device is a substitute device, identification of the corresponding MDD/AIMDD device | MDD/AIMDD Certificate Reference(s) of the devices under MDR application, and the NB Identification |
|---|---|---|---|
| **34342Sauerstoffver3G** | ☒ Class IIa | ☒ N/A | ☒ Certification as follows:<br>Certificate # G1 015692 0502 Rev. 01<br>NB# 0123 |
| **34342SaugAE** | ☒ Class IIa | ☒ N/A | ☒ Certification as follows:<br>Certificate # G1 015692 0502 Rev. 01<br>NB# 0123 |
| **34342SBeutel92** | ☒ Class I devices in sterile condition | ☒ N/A | ☒ Certification as follows:<br>Certificate # G2S 015692 0504 Rev. 01<br>NB# 0123 |
| **34342SetK9H** | ☒ Class I devices in sterile condition | ☒ N/A | ☒ Certification as follows:<br>Certificate # G2S 015692 0504 Rev. 01<br>NB# 0123 |
| **34342SetM9M** | ☒ Class I devices in sterile condition | ☒ N/A | ☒ Certification as follows:<br>Certificate # G2S 015692 0504 Rev. 01<br>NB# 0123 |
| **34342SetS9Z** | ☒ Class I devices in sterile condition | ☒ N/A | ☒ Certification as follows:<br>Certificate # G2S 015692 0504 Rev. 01<br>NB# 0123 |
| **34342SilikonMaskeMY** | ☒ Class IIa | ☒ N/A | ☒ Certification as follows:<br>Certificate # G1 015692 0502 Rev. 01<br>NB# 0123 |
| **34342SkalpelleSL** | ☒ Class IIa | ☒ N/A | ☒ Certification as follows:<br>Certificate # G1 015692 0502 Rev. 01<br>NB# 0123 |
| **34342SkalpellklingeMZ** | ☒ Class IIa | ☒ N/A | ☒ Certification as follows:<br>Certificate # G1 015692 0502 Rev. 01<br>NB# 0123 |
| **34342SMaskeZQ** | ☒ Class IIa | ☒ N/A | ☒ Certification as follows:<br>Certificate # G1 015692 0502 Rev. 01<br>NB# 0123 |
| **34342SondeD9Y** | ☒ Class IIa | ☒ N/A | ☒ Certification as follows:<br>Certificate # G1 015692 0502 Rev. 01<br>NB# 0123 |
| **34342SondeMAJ** | ☒ Class IIa | ☒ N/A | ☒ Certification as follows:<br>Certificate # G1 015692 0502 Rev. 01<br>NB# 0123 |
| **34342SpuelsFA** | ☒ Class IIa | ☒ N/A | ☒ Certification as follows:<br>Certificate # G1 015692 0502 Rev. 01<br>NB# 0123 |
| **34342Stuhldrainage73** | ☒ Class IIa | ☒ N/A | ☒ Certification as follows:<br>Certificate # G1 015692 0502 Rev. 01<br>NB# 0123 |

Heruntergeladen von Gast (IP 216.73.217.10) am 23.11.2026 um 16:21 Uhr · sana.swiss

| Device name or Basic UDI-DI (under MDR application) | MDR Device classification (as proposed by the manufacturer and verified during application review) | If the MDR device is a substitute device, identification of the corresponding MDD/AIMDD device | MDD/AIMDD Certificate Reference(s) of the devices under MDR application, and the NB Identification |
|---|---|---|---|
| **34342SUltraschallK7** | ☒ Class I devices in sterile condition | ☒ N/A | ☒ Certification as follows:<br>Certificate # G2S 015692 0504 Rev. 01<br>NB# 0123 |
| **34342SVerschlussX7** | ☒ Class I devices in sterile condition | ☒ N/A | ☒ Certification as follows:<br>Certificate # G2S 015692 0504 Rev. 01<br>NB# 0123 |
| **34342TKatheterD7** | ☒ Class IIa | ☒ N/A | ☒ Certification as follows:<br>Certificate # G1 015692 0502 Rev. 01<br>NB# 0123 |
| **34342TrachealsQ8** | ☒ Class IIa | ☒ N/A | ☒ Certification as follows:<br>Certificate # G1 015692 0502 Rev. 01<br>NB# 0123 |
| **34342TSBeutelGV** | ☒ Class I devices in sterile condition | ☒ N/A | ☒ Certification as follows:<br>Certificate # G2S 015692 0504 Rev. 01<br>NB# 0123 |
| **34342TSystemNY** | ☒ Class I devices in sterile condition | ☒ N/A | ☒ Certification as follows:<br>Certificate # G2S 015692 0504 Rev. 01<br>NB# 0123 |
| **34342UBeutelA4** | ☒ Class I devices in sterile condition | ☒ N/A | ☒ Certification as follows:<br>Certificate # G2S 015692 0504 Rev. 01<br>NB# 0123 |
| **34342UBeutelMNL** | ☒ Class I devices in sterile condition<br>☒ Class I devices with measuring function | ☒ N/A | ☒ Certification as follows:<br>Certificate #; G2MS 015692 0505 Rev. 00<br>NB# 0123 |
| **34342UrokatheterKD** | ☒ Class IIa | ☒ N/A | ☒ Certification as follows:<br>Certificate # G1 015692 0502 Rev. 01<br>NB# 0123 |
| **34342VBezugSAT** | ☒ Class I devices in sterile condition | ☒ N/A | ☒ Certification as follows:<br>Certificate # G2S 015692 0504 Rev.01<br>NB# 0123 |
| **34342VerneblerSW8** | ☒ Class IIa | ☒ N/A | ☒ Certification as follows:<br>Certificate # G1 015692 0502 Rev. 01<br>NB# 0123 |
| **34342Wendltubus42** | ☒ Class IIa | ☒ N/A | ☒ Certification as follows:<br>Certificate # G1 015692 0502 Rev. 01<br>NB# 0123 |

Heruntergeladen von Gast (IP 216.73.217.10) am 23.11.2026 um 16:21 Uhr · sana.swiss

**Table 2: Devices covered by this letter and for which TÜV SÜD Product Service GmbH is NOT responsible for appropriate surveillance of the corresponding devices under the applicable Directive:**

| Device name or Basic UDI-DI (under MDR application) | MDR Device classification (as proposed by the manufacturer and verified during application review) | If the MDR device is a substitute device, identification of the corresponding MDD/AIMDD device | MDD/AIMDD Certificate Reference(s) of the devices under MDR application, and the NB Identification |
|---|---|---|---|
| **34342LaryngoskopC8** | ☒ Class I reusable surgical instruments | ☒ N/A | ☒ N/A - Device did not require a Notified Body certificate under Directives |

## Confirmation Letter Version History

| Date | TÜV SÜD Product Service GmbH internal reference traceable to each version of the letter | Action |
|---|---|---|
| 2024/09/25 | 713347895, 713347896, 713347898 | Initial issue |

Heruntergeladen von Gast (IP 216.73.217.10) am 23.11.2026 um 16:21 Uhr · sana.swiss