

## EU-DECLARATION OF CONFORMITY

|                                                                        |                                                                            |
|------------------------------------------------------------------------|----------------------------------------------------------------------------|
| Manufacturer                                                           | Wernli AG<br>Henzmannstrasse 18<br>4800 Zofingen<br>Switzerland            |
| Single Registration Number (SRN):<br>Swiss Registration Number (CHRN): | CH-MF-000057741<br>CHRN-MF-20000003                                        |
| EC Representative:                                                     | Wero Swiss Med Kft.<br>Ipartelep utca 6<br>4220 Hajdúböszörmény<br>Hungary |
| Single Registration Number (SRN):                                      | HU-AR-000049642                                                            |

### We declare under our sole responsibility that

the medical device  
with reference no  
with Basic-UDI-DI (product code)

**Wero Swiss Solifix Crepp LF white**  
**see page 2**  
**76300161A9101ZP**

Risk Class

**class 1, unsterile**  
According to Annex VIII, Rule 1

### meets all the provisions of the Regulation (EU) 2017 /745 which apply to it.

|                                                                               |                                                                                                                               |
|-------------------------------------------------------------------------------|-------------------------------------------------------------------------------------------------------------------------------|
| Applied harmonised standards, national standards or other normative documents | EN ISO 10993-1:2021 and applicable parts of the<br>EN ISO 10993 Series<br>EN ISO 15223-1:2021<br>EN ISO 14971:2019 / A11:2021 |
| Conformity assessment procedures                                              | Regulation (EU) 2017 /745<br>according to MDR Article 10 and MDR Article 52<br>(7) Annex I and Annex II                       |
| Place, date                                                                   | Zofingen, 26.10.2026                                                                                                          |

Name and function



.....  
Felix Schoenle, CEO

Wernli AG  
Verbandstoffe-Bandages  
Henzmannstrasse 18  
CH-4800 Zofingen  
[www.weroswiss.com](http://www.weroswiss.com)



Wernli AG  
Verbandstoffe  
Bandages

## Reference Numbers:

Length: 4m stretched

|               |                                             |
|---------------|---------------------------------------------|
| A91010040.040 | Wero Swiss Solifix Crepp LF 4cm x 4m white  |
| A91010060.040 | Wero Swiss Solifix Crepp LF 6cm x 4m white  |
| A91010080.040 | Wero Swiss Solifix Crepp LF 8cm x 4m white  |
| A91010100.040 | Wero Swiss Solifix Crepp LF 10cm x 4m white |
| A91010120.040 | Wero Swiss Solifix Crepp LF 12cm x 4m white |

Length: 20m stretched

|               |                                              |
|---------------|----------------------------------------------|
| A91010040.200 | Wero Swiss Solifix Crepp LF 4cm x 20m white  |
| A91010060.200 | Wero Swiss Solifix Crepp LF 6cm x 20m white  |
| A91010080.200 | Wero Swiss Solifix Crepp LF 8cm x 20m white  |
| A91010100.200 | Wero Swiss Solifix Crepp LF 10cm x 20m white |
| A91010120.200 | Wero Swiss Solifix Crepp LF 12cm x 20m white |