

EU-DECLARATION OF CONFORMITY

Manufacturer	Wernli AG Eggasse 4 4852 Rothrist Switzerland
Single Registration Number (SRN): Swiss Registration Number (CHRN):	CH-MF-000057741 CHRN-MF-20000003
EC Representative:	Wero Swiss Med Kft. Ipartelep utca 6 4220 Hajdúböszörmény Hungary
Single Registration Number (SRN):	HU-AR-000049642

We declare under our sole responsibility that

the medical device with reference no with Basic-UDI-DI (product code)	Wero Swiss Fix see page 2 76300161A091AP
Risk Class	class 1, unsterile According to Annex VIII, Rule 1

meets all the provisions of the Regulation (EU) 2017 /745 which apply to it.

Applied harmonised standards, national standards or other normative documents	EN ISO 10993-1:2021 and applicable parts of the EN ISO 10993 Series EN ISO 15223-1:2021 EN ISO 14971:2019 / A11:2021
Conformity assessment procedures	Regulation (EU) 2017 /745 according to MDR Article 10 and MDR Article 52 (7) Annex I and Annex II
Place, date	Rothrist, 20.08.2026

Name and function



.....
Felix Schoenle, CEO

Reference Numbers:

Length: 4m stretched

A0910020.040	Wero Swiss Fix 2cm x 4m white
A0910040.040	Wero Swiss Fix 4cm x 4m white
A0910060.040	Wero Swiss Fix 6cm x 4m white
A0910080.040	Wero Swiss Fix 8cm x 4m white
A0910100.040	Wero Swiss Fix 10cm x 4m white
A0910120.040	Wero Swiss Fix 12cm x 4m white